

**The Appendix is an integral part of  
Certificate of Accreditation No. 677/2014 of 24/10/2014**

**Accredited entity according to ČSN EN ISO/IEC 17020:2012:**

**Czech Metrology Institute**  
CMI Inspection Body  
Okružní 31, 638 00 Brno

Ordinal number	Inspection field	Inspection type and scope	Procedure identification
1	Inspection of mass flow meters for cryogenic liquids	New and operated devices	Internal Procedure No. 060-MP-C148 (OIML R81, R117), version 14-001
2	Inspection of automatic level gauges	New and operated devices	Internal Procedure No. 060-MP-C504 (OIML R85), version 14-001

**Explanations and abbreviations:**

- OIML R81      Dynamic measuring devices and systems for cryogenic liquids  
OIML R85      Automatic level gauges for measuring the level of liquid in stationary storage tanks  
OIML R117     Dynamic measuring systems for liquids other than water, Part 1: Metrological and technical requirements  
OIML          International Organization of Legal Metrology

